

Statistical Analysis Plan

March 12, 2020

NCT02572856

In Group 1 we will review the results after 10 patients to evaluate the following: general safety of applying the CGM and assessment of any interference with the device in the OR. If the CGM works satisfactorily meeting the above conditions, we will proceed with the remainder of the study. This review has been completed and is safe to continue.

In Group 2 we will review the results after 100 patients to evaluate the following: 1) general safety of applying the CGM, 2) how prevalent the electrocautery interference is and 3) how accurate the measurements are (Mean absolute relative difference - MARD).

A total of ten subjects were enrolled in this study. Statistical analyses were run on these 10 subjects.